



CONBRACO
CALIBRATION SERVICES

A2LA accredited and ISO 9001:2000 Registered

1418 S. Pearl St.
 Pageland, SC 29728
 Tel: (843) 672-1616
 Fax: (843) 672-1625

Repair Questionnaire

If your kit is not functioning correctly or is broken, let us know what the problem is. With this information, we can quickly service and calibrate your kit and return it to you.

Contact Name: _____

Phone: _____

- The pointer doesn't move.
- The pointer movement is jumpy.
- Water is leaking from the:
 - gauge.
 - hoses.
- This kit fails every backflow preventer test.

Comments about your test kit:

Please include this completed form with your gauge. You may also fax it to (843) 672-1625.